

Certificate of analyses

Product: EnzyPlus L-Malic Acid
Art. No: EZA 786
Batch No: 400025
Expiry date: October 2005

Diffchamb certifies that this batch has been approved by the Quality Control Department and that it conforms to the following specifications:

Specific activity

L-MDH \geq 6600 U/ml
GOT \geq 400 U/ml

pH value

Vial 1 9.8 – 10.2

Solubilisation time

Vial 2 \leq 3 min

Performance test

target values \pm 5%

Date: 2004-07-12



Jörgen Plymouth
Quality Manager
Diffchamb AB

Diffchamb AB

F O Petersons Gata 32
SE-421 31 Västra Frölunda
www.diffchamb.com

Telephone +46-(0)31 742 33 50
Telefax +46-(0)31 58 33 70
E-mail diffchamb@diffchamb.se

Org.No. 556314-3303
VAT No. SE 556314330301

Bankgiro 896-5576
Postgiro 85 01 54-6



ISO9001
Certificate
No 11184